

## Description of Data Qualifiers

<u>Qualifier Flag</u>	<u>Description</u>
J	The reported result is an estimate. The value is less than the minimum calibration level but greater than the estimated detection limit (EDL)
U	The analyte was not detected in the sample at the estimated detection limit (EDL)
E	Exceeds calibration range.
D	Dilution data. Result was obtained from the analysis of a dilution
B	Analyte found in sample and associated blank
C	Co-eluting congener
Cxx	Co-elutes with the indicated congener, data is reported under the lowest IUPAC congener. 'Xx' denotes the IUPAC number with the lowest numerical designated congener.
NR	Analyte not reported because of problems in sample preparation or analysis
V	Surrogate recovery is not within method control limits
X	Results from reinjection/repeat/recolumn data
EMPC	Estimated maximum possible concentration. Indicates that a peak is detected but did not meet all met the method required criteria.