



RETURN AND REPAIR FORM

Customer Information

B:

S:

Company Name: Evesham Twp , PD , Marlton , NJDate Received: 06/26/12Date given to service: 06/26/12Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STDProduct: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 WholeSerial Number: AR DC-0066Printer Serial#: ARSim Serial#: DDProbe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

 110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
1815113	FUSE	1	NC/W
MP Labor	Labor	.5	NC/W

Repair Notes: Installed fuse to restore power to unit , ran 50x .1% test , then calibrated unit , then ran 50X .1 % test with no lose of power issues,also left unit off all night, turned unit on next morning no power issue. Unit was also taken apart and completely inspected.

CAL W/QC AND OPS CHECKService Technician JSDate: 06-27-12