

# Return & Repair Form:

## Customer Information:

Company Name: Newton Police Dept. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 3-2-09 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: ARLM - 0198  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARRR</u>
<u>MPC171</u>	<u>CAC</u>	<u>1</u>	<u>N/C WARRR</u>
<u>mPL7BOR</u>		<u>.5</u>	<u>N/C WARRR</u>

Repair Notes:

REPLACED FC / CAC w/ QC & PS CHECKS  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

Service Technician: \_\_\_\_\_

Date: 03/05/2009