

RETURN & REPAIR FORM



CUSTOMER NJ State Police Cust. # \_\_\_\_\_  
 Date ret. 5/25 Carrier UPS Method G1J RMA# \_\_\_\_\_  
 7110 Serial# ARLM 3271 Warranty exp.: \_\_\_\_\_

Reported Problem: \_\_\_\_\_

- Accessories (check all that apply):  Regulator  P.Paper  Scotty V  Printer Rib.  Pouch  
 Sim. Temp. Probe Ser.# \_\_\_\_\_  Keyboard  Organizer  Key  
 Mouthpieces \_\_\_\_\_  Sim. To cuvette hose  Pump to Sim. hose  Power cord  
 Other (specify) \_\_\_\_\_  
 Other (specify) \_\_\_\_\_

Part #	Description	S	Qty	Total
6408455	Fuel Cell		1	
W 12114	SCREW		1	
W 6409514	2 <sup>nd</sup> generation HOSE		1	
W 12117	SCREW		1	
W 6409480	HOSE COVER		1	
W 12118	WASHER		1	
MPCAL 71			1	
	Repair Time		.5	Time \$

TOTAL REPAIR \$ \_\_\_\_\_

REPAIR NOTES:  
 Replaced fuel cell. Upgraded Unit to  
 NJ-3-10. Full QC & OPS CHK. Replaced Breath-  
 HOSE.

20 OCT 04 [Signature]