

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER NJ STATE POLICE / WEST TRENTON Cust # _____

Date of Return: 02/06/2008 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY
OTHER: _____

Product Information:

Product : 7110 7410 6510 6810

Serial # AR LM - 0271

Description: A B PLUS SCREENER

Printer Serial # AR -

Whole Inst. Top Half

Simulator Serial # _____

OTHER : _____

Probe Serial # DD P -

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: ERROR Z3

Part #	Description	Qty	Total
<u>6809512</u>	<u>BARB w/ VALVE</u>	<u>1</u>	<u>74.00</u>
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>98.00</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>44.00</u>

d 216.00
+ S/H/I

Repair Notes: SCREWS LOOSE ON PRINT HEAD / CHECK VALVE IN HOSE BARB
WILL NOT CLOSE - REPLACED / SCREW FOUND UNDER MOTHER BOARD /
SPOT NOTED ON CUVEITE / FC TESTED OK / CAL W/ QC & OPS CHECKS

Technician: [Signature]

Date: 02-08-08