

# RETURN & REPAIR FORM



CUSTOMER NJSP Cust. # \_\_\_\_\_

Date ret. 12/20 Carrier UPS Method GROUND RMA# \_\_\_\_\_

7110 Serial# ARLM-027Z Warranty exp.: \_\_\_\_\_

Reported Problem: UPGRADE FC - FIRMWARE

- Accessories (check all that apply):  Regulator  P.Paper  Scotty V  Printer Rib.  Pouch  
 Sim. Temp. Probe Ser.# \_\_\_\_\_  Keyboard  Organizer  Key  
 Mouthpieces \_\_\_\_\_  Sim. To cuvette hose  Pump to Sim. hose  Power cord  
 Other (specify) \_\_\_\_\_  
 Other (specify) \_\_\_\_\_

<u>Part#</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
6808455	Fuel Cell		1	
MPCAL71			1	
		Repair Time	.5	Time \$

TOTAL REPAIR \$ \_\_\_\_\_

REPAIR NOTES:  
Replaced fuel cell, Upgraded firmware, Calibrated  
Unit.

Date 31 Dec 07 Technician [Signature]

Every above checked item is included  Shipping \_\_\_\_\_