

# Dräger safety

## Return & Repair Form

**Customer Information:**

CUSTOMER NJ state POLICE Cust # \_\_\_\_\_  
 Date of Return: 10/12 Carrier: FedEx Method: 2nd

**Product Information:**

Product : 7110 7410 6510 6810 Serial # ARLM 0272  
 Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
**Accessories:** (check all that apply):  
 110VAC Adapter       Regulator       Mag Card Rdr # \_\_\_\_\_  
 Printer Paper       Printer Ribbon       Casio # \_\_\_\_\_  
 Mouthpieces       Carrying Case       Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

**Repair Information:**

Test # \_\_\_\_\_

Reason for Return: \_\_\_\_\_

Part #	Description	Qty	Total
<u>W 6309514</u>	<u>BREATH HOSE</u>	<u>1</u>	<u>N/C</u>
<u>W 12114</u>	<u>Screw 3mm x</u>	<u>1</u>	<u>N/C</u>
<u>W 12117</u>	<u>SCREW 3mm x</u>	<u>1</u>	<u>N/C</u>
<u>W 12118</u>	<u>WASHER</u>	<u>1</u>	<u>N/C</u>
<u>W MPAL71</u>	<u>7110 CAL</u>	<u>1</u>	<u>N/C</u>

Repair Notes: REPLACED BREATH HOSE CAL OPS CHECK + O.C.

Technician: Carl Dely

Date: 10/24/05