

## Return & Repair Form

### Customer Information:

CUSTOMER NJ STATE POLICE / WEST TRENTON Cust # \_\_\_\_\_  
 Date of Return: 02/06/2008 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY  
 OTHER: \_\_\_\_\_

### Product Information:

Product : 7110 7410 6510 6810 Serial # AR LM - 0272  
 Description: A B PLUS SCREENER Printer Serial # AR -  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # DD P -  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: HIGH IR READINGS

Part #	Description	Qty	Total
<u>MPCALY1</u>	<u>7110 CAC</u>	<u>1</u>	<u>98<sup>00</sup></u>
<u>6808455</u>	<u>FE</u>	<u>1</u>	<u>192<sup>00</sup></u>
<u>MPLAB01C</u>	<u>LABOR</u>	<u>.5</u>	<u>44<sup>00</sup></u>

334<sup>00</sup>  
+ 514<sup>00</sup>

Repair Notes: CUVETTE INSPECTION LOOKS OK / CAC W/ QC'S OPS CHECKS + ADDITIONAL .10% TESTING - FC FAILED  
QC - REPLACED - RE-CAC + QC'S OPS CHECKS  
HIGH IR READINGS NOT DUPLICATED

Technician: [Signature]

Date: 02-07-08