

RETURN AND REPAIR FORM

Customer Information

B:

S:

Company Name: ADTV - New Jersey SP

Date Received: 8-8-11 Date given to service: _____

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR LM - 0272

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:


Test#:

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>N/CW</u>
<u>6809500</u>	<u>CUvette</u>	<u>1</u>	<u>N/CW</u>
<u>mp Cal 71</u>	<u>Cal.</u>	<u>1</u>	<u>N/CW</u>
<u>6809512</u>	<u>Valve</u>	<u>1</u>	<u>N/CW</u>
<u>13336</u>	<u>Hose</u>	<u>22IN</u>	<u>N/CW</u>
<u>mp labor</u>	<u>labor</u>	<u>.5</u>	<u>N/CW</u>

Repair Notes: New Fuel Cell # ARCA-0665
Install new Hoses.

Cal w/ DC steps check
Unit had spots on cuvette mirrors that could
not be removed, installed new cuvette

Service Technician _____



Date: 8-11-11