


RETURN & REPAIR FORM



CUSTOMER NJ State Police Cust. # _____

Date ret. 5/25 Carrier UPS Method G1D RMA# _____

7110 Serial# ARLM 0283 Warranty exp.: _____

Reported Problem: _____

- Accessories (check all that apply):
- Regulator
 - P.Paper
 - Scotty V
 - Printer Rib.
 - Pouch
 - Sim. Temp. Probe Ser.# _____
 - Keyboard
 - Organizer
 - Key
 - Mouthpieces _____
 - Sim. To cuvette hose
 - Pump to Sim. hose
 - Power cord
 - Other (specify) _____
 - Other (specify) _____

Part#	Description	\$	Qty	Total
W 6808455	Fuel Cell		1	
W 12114	Screw		1	
W 6809514	2 nd Generation Hose		1	
W 12117	Screw		1	
W 6809480	HOSE COVER		1	
W 12118	WASHER.		1	
MPCAL41			1	
	Repair Time		.5	Time \$

TOTAL REPAIR \$ _____

REPAIR NOTES:
 Replaced Breath-hose, Replaced Fuel Cell. Full QC & OPS CHE WITH CALIBRATION. Upgraded UNIT TO NJ-3.10.

Date 20 OCT 04 Technician [Signature]
 Government-owned item is included Shipping