

# Return & Repair Form:

## Customer Information:

Company Name: New Jersey State Police B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 1-16-09 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - (LPS) - DHL - USPS Method: (GRD) - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 - (7110) - 6510 - 6810 Serial #: AR LM-0283  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: STUCK ON "MEMORY TEST" AT START UP

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	
<u>LABOR</u>		<u>.5</u>	

## Repair Notes:

PROBLEM NOT DUPLICATED - MEMORY @ 31.3% / CLEARED MEMORY  
FC IS UF - 06/2004 / REPLACED  
CAL W/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 01-20-09