RETURN & REPAIR FORM

ALCOTEST 7110

CUSTOMER Dept. / NJ State police Cust. # 1113

Date ret. 10/12 Carrier Fed Exp Method ECON RMA# 15652

7110 Serial# ARLM 0284 Warranty exp.: —

Reported Problem: See attached note

Accessories (check all that apply): ☐ Regulator ☐ P.Paper ☐ Scotty V ☐ PrinterRib. ☐ Pouch
Sim.Temp.Probe Ser.# _____ ☐ Keyboard ☐ Organizer ☐ Key
☐ Mouthpieces_____ ☐ Sim. to cuvette hose ☐ Pump to Sim. hose ☐ Power cord
☐ Other (specify) _____
☐ Other (specify) _____

Part#	Description	$	Qty	Total
		Repair Time		Time $

TOTAL REPAIR $ _____

REPAIR NOTES:
DID 10 TESTS OK - CHECK FLOW SENSOR - OK - ADJ FLOW SENSOR VOLTAGE TO MAX 2.08 VOLTS - FLOW CHECK 1000 = 8.6 LPM - AA = 1102 = 13.6 LPM / MA 1112 = 14.0 LPM

RECHECK FLOW 10-20- 1112 = 15.0 ADJ 1105 = 14.0 LPM

I NEVER COULD DUPLICATE PROBLEM - FLOW ADJUSTMENT SEEMS TO HAVE TAKE CARE OF COMPLAINT.

Date 10-19-98 Technician [signature]
Every above checked item is included ☐ Shipping _____

G:\HOME\COMMON\FORMS\RRFORM.WPD