

RETURN & REPAIR FORM



CUSTOMER NJ State Police Cust. # \_\_\_\_\_  
 Date ret. 5/25 Carrier UPS Method G1D RMA# \_\_\_\_\_  
 7110 Serial# ARLM0284 Warranty exp.: \_\_\_\_\_

Reported Problem: \_\_\_\_\_

- Accessories (check all that apply):  Regulator  P.Paper  Scotty V  Printer Rib.  Pouch  
 Sim. Temp. Probe Ser.# \_\_\_\_\_  Keyboard  Organizer  Key  
 Mouthpieces \_\_\_\_\_  Sim. To cuvette hose  Pump to Sim. hose  Power cord  
 Other (specify) \_\_\_\_\_  
 Other (specify) \_\_\_\_\_

Part#	Description	\$	Qty	Total
6808455	Fuel Cell		1	
12114	Screw		1	
6809514	2 <sup>nd</sup> Generation Hose		1	
12117	Screw		1	
6809480	Hose Cover		1	
12118	WASHER		1	
MPCAL H			1	
			Repair Time	Time \$
				.5

TOTAL REPAIR \$ \_\_\_\_\_

REPAIR NOTES:  
 Replaced fuel cell. Replaced Breath-hose. Full  
 PE & OPS CHK WITH CALIBRATION. Upgraded D.O.T  
 TO N.J-3.10.

Date 20 OCT 04 Technician [Signature]  
 Approved by: \_\_\_\_\_