

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER NJ STATE POLICE Cust # _____

Date of Return: 10/12 Carrier: FedEx Method: 2ND

Product Information:

Product: 7110 7410 6510 6810 Serial # ARCM0284

Description: A B PLUS SCREENER _ Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: ERROR 112

Part #	Description	Qty	Total
<u>W MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	<u>N/C</u>

Repair Notes: CLEARED ERROR 112, CAL OPS CHECK + Q.C.

Technician: Art Dly

Date: 10/24/05