RETURN & REPAIR FORM

ALCOTEST 7110

CUSTOMER N J Cust. #

Date ret. 1/16/01 Carrier Fed Exp Method 3 Day RMA#

7110 Serial# ARLM 0287 Warranty exp.:

Reported Problem:

Accessories (check all that apply): ☐ Regulator ☐ P.Paper ☐ Scotty V ☐ PrinterRib. ☐ Pouch
Sim.Temp.Probe Ser.# ☐ Keyboard ☐ Organizer ☐ Key
☐ Mouthpieces ☐ Sim. to cuvette hose ☐ Pump to Sim. hose ☐ Power cord
☐ Other (specify)
☐ Other (specify)

Part#	Description	$	Qty	Total
		Repair Time		Time $

TOTAL REPAIR $

REPAIR NOTES:
Replaced Small Pre-Warming Track With the New large Style Pre-Warming Track. Upgraded NJ-3.2 TO NJ-3.4. Full QC & OPS CHK. Unit Had TO Be Calibrated.

Date 01/17/01 Technician EW

Every above checked item is included ☐ Shipping

G:\HOME\COMMON\FORMS\RRFORM.WPD