

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER NJ State Police Cust # _____
 Date of Return: 5/23 Carrier: FedEx Method: 2 Day

Product Information:

Product: 7110 7410 6510 6810 Serial # ARLM 0287
 Description: A B PLUS SCREENER _____ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: EC Reading high

Part #	Description	Qty	Total
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	
<u>12008</u>	<u>1/2" Recess Bumper</u>	<u>1</u>	

Repair Notes: EC-Sensor was too high. Calibrated unit to adjust EC correctly. Full QC & ops check
Replaced missing bumper.

Technician: CAF

Date: 5/25/06