


RETURN & REPAIR FORM



CUSTOMER NJ State Police Cust. # \_\_\_\_\_

Date ret. 5/25 Carrier UPS Method 611 RMA# \_\_\_\_\_

7110 Serial# ARLM 0287 Warranty exp.: \_\_\_\_\_

Reported Problem: \_\_\_\_\_

- Accessories (check all that apply):  Regulator  P.Paper  Scotty V  Printer Rib.  Pouch  
 Sim. Temp. Probe Ser.# \_\_\_\_\_  Keyboard  Organizer  Key  
 Mouthpieces \_\_\_\_\_  Sim. To cuvette hose  Pump to Sim. hose  Power cord  
 Other (specify) \_\_\_\_\_  
 Other (specify) \_\_\_\_\_

Part#	Description	\$	Qty	Total
6808455	Fuel Cell		1	
W 12114	SCREW		1	
W 6809514	BREATH-HOSE		1	
W 12117	SCREW		1	
W 6409480	HOSE COVER		1	
W 12118	WASHER		1	
MPCAL 71			1	
	Repair Time		.5	Time \$

TOTAL REPAIR \$ \_\_\_\_\_

REPAIR NOTES:

Replaced Fuel Cell, Replaced Breath-Hose, Upgraded  
 FIRMWARE TO NJ-3-10. Full DC & OPS CHK WITH  
 CALIBRATION.

DATE OF REPAIR 26 OCT 04 Technician [Signature]

Customer Name & Address \_\_\_\_\_