

# Return & Repair Form

## Customer Information:

Company Name: NJSP B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 3.24.09 Date Given to Service: 3.25.09  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR NK - 0017  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: FUEL CCI

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL 71</u>	<u>CAC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>		<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

REPLACED FC  
CAL W/ QC & OPS CHECKS

Service Technician: [Signature]

Created on 4/10/2008 7:22 AM

Date: 03.25.09