

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Logan PD Cust # 53658
 Date of Return: 5/23 Carrier: FedEx Method: 2011

Product Information:

Product: 7110 7410 6510 6810 Serial # ARNK-00'8
 Description: A B PLUS SCREENER _ Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: IR + EC reading Low

Part #	Description	Qty	Total

Repair Notes: Confirmed unit running low. Returned unit w/o cal. Trooper Bill Cross will cal on site per their process

Technician: CPD

Date: 5/30/06