

RETURN AND REPAIR FORM



Customer Information

Company Name: Logan TWP P.D. Bridgeport, N.J.

Date Received: 10/14/11 Date given to service: _____

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810 7110 8610 7510 DT5000 7410 Upper-half 9510 7410 Whole

Serial Number: AR NK-0018
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
<u>16808455</u>	<u>Fuel Cell</u>	<u>1</u>	<u>N/C W</u>
<u>mp cal 71</u>	<u>Calibrate</u>	<u>1</u>	<u>N/C W</u>
<u>mp labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C W</u>

Repair Notes: Replaced Fuel Cell ARAC-0510 with Fuel Cell ADEF-D651, because of Fuel cell leakage. Replaced capacitor located at position C-167 to restore modern connection.

Cal w/dc; ops check

Service Technician: [Signature] Date: 10-17-11