RETURN & REPAIR FORM

ALCOTEST 7110

CUSTOMER Dover Police Cust. #

Date ret. 4/18 Carrier Fed Exp Method Priority RMA#

7110 Serial# ARNK-0028 Warranty exp.:

Reported Problem:

Accessories (check all that apply): ☐ Regulator ☐ P.Paper ☐ Scotty V ☐ Printer Rib. ☐ Pouch
Sim. Temp. Probe Ser.# ☐ Keyboard ☐ Organizer ☐ Key
☐ Mouthpieces ☐ Sim. To cuvette hose ☐ Pump to Sim. hose ☐ Power cord
☐Other (specify)
☐Other (specify)

Part#	Description	$	Qty	Total
		Repair Time		Time $

TOTAL REPAIR $

REPAIR NOTES:
Calibration Value for the EC-Sensor was maxed out. Moved Calibration Value down to the Right Value and Calibrated Unit, full QC & OPS CHK.

Date 19 APR 05 Technician

Every above checked item is included ☐ Shipping

G:\HOME\COMMON\FORMS\RRFORM.WPD