RETURN & REPAIR FORM

CUSTOMER Dover Police Cust. #

Date ret. 4/18 Carrier Fed Ex Method Priority RMA#

7110 Serial# ARNK-0028 Warranty exp.:

Reported Problem:

Accessories (check all that apply): ☐ Regulator ☐ P.Paper ☐ Scotty V ☐ Printer Rib. ☐ Pouch
Sim. Temp. Probe Ser.# ☐ Keyboard ☐ Organizer ☐ Key
☐ Mouthpieces ☐ Sim. To cuvette hose ☐ Pump to Sim. hose ☐ Power cord
☐Other (specify)
☐Other (specify)

Part#	*Description*	*$*	*Qty*	*Total*
		Repair Time		Time $

TOTAL REPAIR $

REPAIR NOTES:
Calibration Value for the EC-Sensor was marked Out. Moved Calibration Value down to the Right Value and Calibrated Unit, Full QC & OPS CHK.

Date 19 APR 05 Technician

Every above checked item is included ☐ Shipping