

# Return & Repair Form

## Customer Information:

Company Name: Passipany P.D. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 9-3-08 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY 3DY Other \_\_\_\_\_

## Product Information:

Product: 7410 7110 - 6510 - 6810 Serial #: AR NK-0031  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>680745B</u>	<u>DETECTOR</u>	<u>1</u>	<u>N/C WARRR</u>
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARRR</u>
<u>MPLABOR</u>		<u>.5</u>	<u>N/C WARRR</u>

### Repair Notes:

DEBRIS IN CUVETTE - BLEW OUT W/ AIR / REPLACED DETECTOR AS P.M. /  
IR FILTER LOOSE - SECURED IN PLACE. / FC 15 (4E) INSTALLED 3-17-08  
CAL W/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 09.10.08