

Return & Repair Form:

Customer Information:

Company Name: Little Egg Harbor B: _____ S: _____
 Date Received: 4-13-10 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6810-7510 Serial #: AR NK-0032
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC

Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C w</u>
<u>6808486</u>	<u>Plate</u>	<u>1</u>	<u>N/C w</u>
<u>MP cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C w</u>
<u>12003</u>	<u>Hose band</u>	<u>1</u>	<u>N/C w</u>
<u>MP Labor</u>	<u>Labor</u>	<u>1.5</u>	<u>N/C w</u>

Repair Notes:

New FC # ARAN-1016
Cal w/DC! Ops check

Service Technician: JFS

Date: 04/14/2010