

## Return & Repair Form

### Customer Information:

CUSTOMER CRANBURY TWP. P.D. Cust # 52291/52292

Date of Return: 11/16/2007 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY  
OTHER: STO CEMENTITE

### Product Information:

Product 7110 7410 6510 6810 Serial # ARNIK - 0033

Description: A B PLUS SCREENER Printer Serial # AR -

Whole Inst. Top Half Simulator Serial # \_\_\_\_\_

OTHER: \_\_\_\_\_ Probe Serial # DD P -

Demo Unit  Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # \_\_\_\_\_
- Printer Paper
- Printer Ribbon
- Casio # \_\_\_\_\_
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): \_\_\_\_\_

Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: FC RESULTS = 0.000

Part #	Description	Qty	Total
<u>8100902</u>	<u>FLOW SENSOR</u>	<u>1</u>	<u>WARR N/C</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>WARR N/C</u>
<u>6808486</u>	<u>PLATE</u>	<u>1</u>	<u>WARR N/C</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>1</u>	<u>WARR N/C</u>

Repair Notes: ERROR 41 = FLOW SENSOR BROKEN / FC CORROSION - REPLACE BOTH. CAL W/ QC & OPS CHECKS

Technician: [Signature]

Date: 11-19-07