

Return & Repair Form:

Customer Information:

Company Name: Don Pomlot B: _____ S: _____
 Date Received: 4-9-10 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR NIK - 0035
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: Print Cable

| Part Number | Description | Qty | Total Cost |
|----------------|--------------|-----------|--------------|
| <u>8314305</u> | <u>Cable</u> | <u>1</u> | <u>N/C W</u> |
| <u>MPLC600</u> | <u>Labor</u> | <u>.5</u> | <u>N/C W</u> |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |
| | | | |

Repair Notes:

Install new printer cable, performed print test, Diagnostic test, and STND-check x2.

Service Technician: [Signature]

Date: 4/9/2010