

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Middlesex Boro Police Dept. Cust # _____

Date of Return: 2/19/17 Carrier: Fed Ex Method: 2-Day

Product Information:

Product : 7110 7410 6510 6810 Serial # ARNK-0037

Description: A B PLUS SCREENER Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER : _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: repair per attached EC Problem

Part #	Description	Qty	Total
<u>4808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>6809512</u>	<u>Hose barb w/check valve</u>	<u>1</u>	
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPLAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced leaky FLC and sticky check valve. Full cal ops check + Q.S.

Technician: CRD

Date: 2/20/07