# RETURN & REPAIR FORM

ALCOTEST 7110

CUSTOMER Stamford police Cust. #

Date ret. 4/11 Carrier Fed Exp Method Priority RMA#

7110 Serial# ARNK-0039 Warranty exp.:

Reported Problem: Ambient Air Check Error

Accessories (check all that apply): ☐ Regulator ☐ P.Paper ☐ Scotty V ☐ Printer Rib. ☐ Pouch
Sim. Temp. Probe Ser.# ☐ Keyboard ☐ Organizer ☐ Key
☐ Mouthpieces ☐ Sim. To cuvette hose ☐ Pump to Sim. hose ☐ Power cord
☐Other (specify)
☐Other (specify)

| Part# | Description | $ | Qty | Total |
|---|---|---|---|---|
| W/ | misc part | | 1 | |
| | | | | |
| | | | | |
| | | | | |
| W/ | MPCAL71 | | 1 | |
| | | Repair Time | | Time $ |

TOTAL REPAIR $

REPAIR NOTES:
Replaced Cuvette Due to Ambient Air CHK failure. Solution found in Cuvette. Replaced Cuvette from the 7110 Damaged in Shipping.

Date 11 APR 05 Technician

Every above checked item is included ☐ Shipping