

RETURN AND REPAIR FORM



Customer Information

B:

S:

Company Name: Stratford Borough P.D. , Stratford,NJ

Date Received: 08/03/12

Date given to service: 08/03/12

Carrier: FedEx UPS USPS

Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR NK-0039

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	.5	NC/W

Repair Notes: Unit need recalibration , recalibrated unit.

CAL W/QC AND OPS CHECK

Service Technician JS

Date: 08-03-12