

Return & Repair Form

Customer Information:

CUSTOMER City of Burlington Police Dept Cust # _____
 Date of Return: 1/22/7 Carrier: UPS Method: 2-Day

Product Information:

Product: 7110 7410 6510 6810 Serial # ARNK-0040
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____ ** not packaged well*
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: repair per attached Erratic IR REC

Part #	Description	Qty	Total
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Could not determine any problems with this unit. I suspect a seal leak is the cause of erratic results. Full cal ops check and Q.C.

Technician: CRD

Date: 1/26/07