

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Burlington City Cust # _____
 Date of Return: 4-18-7 Carrier: UPS Method: Ground

Product Information:

Product : (7110) 7410 6510 6810 Serial # ARNK-0040
 Description: A B PLUS SCREENER Printer Serial # _____
 Whole Inst. Top Half Simulator Serial # _____
 OTHER : _____ Probe Serial # _____
 Demo Unit Returned to stock on: _____ *** NOT PACKAGED well**
 Accessories: (check all that apply):
 110VAC Adapter Regulator Mag Card Rdr # _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (specify): _____
 Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: EC + IR problems

Part #	Description	Qty	Total
8314685	Case, bottom w/lock	1	691.00
12064	nipple, repair top cover	1	33.00
MPLABOR	LABOR	1	78.00
W 6808455	Fuel Cell	1	N/C
W MPLABOR	LABOR	.5	N/C
W MPLACT1	7110 CAL	1	N/C

Repair Notes: Replaced bottom case, top cover nipple due to shipping damage. Replaced questionable F/C. Full cal ops check + O.C.

Technician: CRD

Date: 6/19/07