

Return & Repair Form

RETURN
FX-STD

Customer Information:

CUSTOMER Burlington PD Cust # _____
 Date of Return: 11/2/2008 Carrier: FX UPS DHL USPS Method: GRD NDA 2DY 3DY
 OTHER: STD

Product Information:

Product: 7110 7410 6510 6810 Serial # AR NK - 0090
 Description: A B PLUS SCREENER Printer Serial # AR -
 Whole Inst. Top Half Simulator Serial # _____
 OTHER: _____ Probe Serial # DDUJPZ-010

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter
- Regulator
- Mag Card Rdr # _____
- Printer Paper
- Printer Ribbon
- Casio # _____
- Mouthpieces
- Carrying Case
- Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: ~~ERR~~ DROP IN IR/EC RESULTS

Part #	Description	Qty	Total
<u>MPCAL 71</u>	<u>CAL</u>	<u>1</u>	<u>MCWARR</u>
<u>12025</u>	<u>PROBE</u>	<u>1</u>	<u>WARR</u>

Repair Notes: EC WAS INSTALLED 6-19-07 (YA) / CAL W/ QC EOPS ✓
PROBLEM NOT REPLICATED

PROBE DDUJPZ-010 FAILED - RETURN TO BLOCK STOCK
REPLACED W/ DDUJPZ-254

Technician: [Signature]

Date: 01-15-08