

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Burlington City PD

Date Received: 9-22-11 Date given to service: \_\_\_\_\_

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR NK - 0040  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
6808455	Fuel Cell	1	N/C W
6808486	Plate	1	N/C W
mp cal 71	Calibration	1	N/C W
mp labor	Labor	1.5	N/C W

Repair Notes: Fuel Cell AR YA-0399 replaced because of leakage, replaced with Fuel Cell ARCH-0450

Cal w/DC ops check

Service Technician: [Signature] Date: 9-27-11