

RETURN AND REPAIR FORM



Customer Information

B: _____

S: _____

Company Name: Burlington City P.D., Burlington, NJ

Date Received: 01/25/2017

Date given to service: 01/25/2017

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810 7110 7510 7410 Upper-half 7410 Whole
 8610 DT5000 9510

Serial Number: AR NK-0040
 Printer Serial#: AR
 Sim Serial#: DD
 Probe Serial#: DD

Warranty Expires: _____

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#: _____

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	

Repair Notes: Replaced defective fuel cell.

Fuel Cell = ARJL-0336

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 02/13/2017