

1  
NJ

# Dräger safety

Internal Use Only ✓

## Return & Repair Form

### Customer Information:

CUSTOMER Rutgers State Univ Cust # 53078/53079  
 Date of Return: 7/23/07 Carrier: Fedex Method: P1

### Product Information:

Product: 7110 7410 6510 6810 Serial # ARNK-0042  
 Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): \_\_\_\_\_  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: FC PROB

Part #	Description	Qty	Total
<u>6808486</u>	<u>FUEL CELL PLATE</u>	<u>1</u>	<u>WARR</u>
<u>6808455</u>	<u>FUEL CELL</u>	<u>1</u>	<u>WARR</u>
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	<u>WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>WARR</u>

Repair Notes: REPLACE FC & PLATE - FULL CAL & QC

Technician: [Signature]

Date: 07.24.07