

Return & Repair Form

Customer Information:

Company Name: Manalapan TWP B: _____ S: _____
 Date Received: 9-29-08 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410 - 110 - 6510 - 6810 Serial #: AR NK - 0045
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top ½ Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return:			
Part Number	Description	Qty	Total Cost
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPCAL71</u>	<u>CAL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MPLABOR</u>	<u>LABOR</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

REPLACED FC - UNIT WILL NOT TURN ON - PROBLEM HAPPENED WHEN
COORDINATOR ATTEMPTED UPDATE - HAD TO BOOT STRAP FIRMWARE.
CAL W/ QC & OPS CHECKS

LOADED VERSION VIA DIRECT & ALSO TO TEST COMMUNICATION PORT - TEST = OK

Service Technician: [Signature]

Date: 10-02-08