

Dräger safety

Return & Repair Form

Customer Information:

CUSTOMER Boro of South Plainfield. Cust # _____

Date of Return: 5-11-7 Carrier: FedEx Method: 2-Day

Product Information:

Product: 7110 7410 6510 6810 Serial # ARNK-0056

Description: A B PLUS SCREENER Printer Serial # _____

Whole Inst. Top Half Simulator Serial # _____

OTHER: _____ Probe Serial # _____

Demo Unit Returned to stock on: _____

Accessories: (check all that apply):

- 110VAC Adapter Regulator Mag Card Rdr # _____
- Printer Paper Printer Ribbon Casio # _____
- Mouthpieces Carrying Case Dry Gas
- Other (specify): _____

Warranty Expiration Date: _____

Repair Information:

Test # _____

Reason for Return: EC Problems

Part #	Description	Qty	Total
<u>6808455</u>	<u>Fuel Cell</u>	<u>1</u>	
<u>MP LABOR</u>	<u>LABOR</u>	<u>.5</u>	
<u>MPCAL71</u>	<u>7110 CAL</u>	<u>1</u>	

Repair Notes: Replaced leaky FLC. Full cal ops check + QC.

Technician: CRD

Date: 5/11/07