

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER Alpine Police Cust # 081105  
 Date of Return: 8/18 Carrier: Fed Exp Method: STK

### Product Information:

Product : 7110 7410 6510 Printer Simulator Probe Serial # ARNK-00617  
 Description: A B PLUS SCREENER \_ Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Serial # \_\_\_\_\_  
 OTHER : \_\_\_\_\_ Probe Serial # \_\_\_\_\_

Demo Unit  Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

- 110VAC Adapter       Regulator       Mag Card Rdr # \_\_\_\_\_
- Printer Paper       Printer Ribbon       Casio # \_\_\_\_\_
- Mouthpieces       Carrying Case       Dry Gas

Other (specify): power strip

Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: \_\_\_\_\_

W = WARRANTY  
C = PER CONTRACT

Part #	Description	Qty	Total
<u>W 12016</u>	<u>SOLENOID NIPPLE</u>	<u>1</u>	
<u>W 13336</u>	<u>VITON TUBING 4"</u>	<u>1</u>	<u>PER CONTRACT</u>
<u>W 13336</u>	<u>VITON TUBING 9"</u>	<u>1</u>	
<u>W 6808455</u>	<u>FUEL CELL</u>	<u>1</u>	
<u>W 12029</u>	<u>Power Supply</u>	<u>1</u>	
<u>C MPCAL71</u>	<u>CAL</u>	<u>1</u>	
<u>C MPL1900A</u>	<u>LABOR</u>	<u>.5</u>	
<u>W 6809456</u>	<u>IR DETECTOR</u>	<u>1</u>	

Repair Notes: BROUGHT 7110 UP TO DATE TO LIKE NEW CONDITION PER CONTRACT. REPLACED POWER SUPPLY + IR DETECTOR UNDER WARRANTY. CAL, SC + OPS CHECK

Technician: Outdy / E

Date: 8/22/05