

# RETURN & REPAIR FORM




CUSTOMER NJ / S. BAUNSWICK Cust. # \_\_\_\_\_

Date ret. 6/1 Carrier Fed Exp Method GRD RMA# \_\_\_\_\_

7110 Serial# AKNK0074 Warranty exp.: \_\_\_\_\_

Reported Problem: \_\_\_\_\_  
 \_\_\_\_\_

- Accessories (check all that apply):  Regulator  P.Paper  Scotty V  Printer Rib.  Pouch  
 Sim. Temp. Probe Ser.# \_\_\_\_\_  Keyboard  Organizer  Key  
 Mouthpieces \_\_\_\_\_  Sim. To cuvette hose  Pump to Sim. hose  Power cord  
 Other (specify) \_\_\_\_\_  
 Other (specify) \_\_\_\_\_

<u>Part#</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
	<u>Fuel Cell</u>		<u>1</u>	
<u>MPAL 71</u>			<u>1</u>	
		Repair Time		Time \$

TOTAL REPAIR \$ \_\_\_\_\_

REPAIR NOTES:

Calibrated Unit. Replaced fuel Cell. Full Op  
& OPS CHIL. Upgraded UNIT TO NJ 3-104

Date 19 Jul 04 Technician [Signature]

Every above checked item is included  Shipping \_\_\_\_\_