

# Dräger safety

## Return & Repair Form

### Customer Information:

CUSTOMER PARK RIDGE PD Cust # \_\_\_\_\_  
 Date of Return: 8/30 Carrier: Fed Ex Method: GLD

### Product Information:

Product: (7110) 7410 6510 6810 Serial # ARNK-0075  
 Description: A B PLUS SCREENER \_\_\_\_\_ Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 OTHER: \_\_\_\_\_ Probe Serial # \_\_\_\_\_  
 Demo Unit  Returned to stock on: \_\_\_\_\_  
 Accessories: (check all that apply):  
 110VAC Adapter  Regulator  Mag Card Rdr # \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (specify): Power Bio(1) surge prot., phone cable  
 Warranty Expiration Date: \_\_\_\_\_

### Repair Information:

Test # \_\_\_\_\_

Reason for Return: fuel cell upgrade

Part #	Description	Qty	Total
6808455	FUEL CELL	1	
MPCAL 71	7110 CAL	1	

Repair Notes: REPLACED FUEL CELL PER UPGRADE AGREEMENT.  
CAL, QC + OPS CHECK

Technician: Carl Dely

Date: 8/31/05