

# RETURN & REPAIR FORM



CUSTOMER HAWTHORNE POLICE Cust. # \_\_\_\_\_

Date ret. 9/7 Carrier FedEx Method GT RMA# \_\_\_\_\_

7110 Serial# ACNL-0036 Warranty exp.: \_\_\_\_\_

Reported Problem: See letter attached - fuel cell - cal

- Accessories (check all that apply):
- Regulator
  - P.Paper
  - Scotty V
  - Printer Rib.
  - Pouch
  - Sim. Temp. Probe Ser.# \_\_\_\_\_
  - Keyboard
  - Organizer
  - Key
  - Mouthpieces \_\_\_\_\_
  - Sim. To cuvette hose
  - Pump to Sim. hose
  - Power cord
  - Other (specify) \_\_\_\_\_
  - Other (specify) \_\_\_\_\_

Part#	Description	\$	Qty	Total
w/ 6808455	Fuel Cell		1	
w/ MPCAL71			1	
		Repair Time	.5	Time \$

TOTAL REPAIR \$ \_\_\_\_\_

REPAIR NOTES:  
Replaced fuel cell. Calibrated Unit with  
full GC & OBS CHA. Upgraded Unit to NJ-SP.

Date 08 SEPT 04 Technician [Signature]

Every above checked item is included  Shipping \_\_\_\_\_