

# Return & Repair Form:

## Customer Information:

Company Name: HAUWORTH P.D. NJ. B: \_\_\_\_\_ S: \_\_\_\_\_  
 Address where rec'd from: \_\_\_\_\_  
 Date Received: 11-14-08 Date Given to Service: 11-15-08  
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410 - 7110 - 6510 - 6810 Serial #: AR NL - 0036  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR - \_\_\_\_\_  
 Whole Instrument Top ½ Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P - \_\_\_\_\_

**ACCESSORIES:**

110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

Test #

Reason for Return: HIGH IR RESULTS

Part Number	Description	Qty	Total Cost
<u>6809500</u>	<u>CUVETTE</u>	<u>1</u>	<u>N/C W/WRK</u>
<u>6808455</u>	<u>FC</u>	<u>1</u>	<u>N/C W/WRK</u>
<u>MPCIR 71</u>	<u>CAC</u>	<u>1</u>	<u>N/C W/WRK</u>
<u>MPLABOR</u>		<u>1 HR</u>	<u>N/C W/WRK</u>

## Repair Notes:

CUVETTE HAD DEFECT ON MIRROR - REPLACED CUVETTE  
REPLACE FC  
CAL w/ QC & OPS CHECKS

Service Technician: [Signature]

Date: 11-18-2008