

Return & Repair Form:

Customer Information:

Company Name: Hawthorne PD B: _____ S: _____
 Date Received: 10/21/09 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR NL - 0036
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: _____

Part Number	Description	Qty	Total Cost
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C WARR</u>

Repair Notes:

Inspect unit including cleaning CUVETTE, also inspect Filter.

Cal w/DC; Ops check

Service Technician: js

Date: 10/21/09