

RETURN & REPAIR FORM



CUSTOMER NE State Police Cusl. # _____
70/200 G12

Date ret. 5/25 Carrier UPS Method G12 RMA# _____

7110 Serial# ARNL-0046 Warranty exp.: _____

Reported Problem: _____

- Accessories (check all that apply): Regulator P.Paper Scotty V Printer Rib. Pouch
 Sim. Temp. Probe Ser.# _____ Keyboard Organizer Key
 Mouthpieces _____ Sim. To cuvette hose Pump to Sim. hose Power cord
 Other (specify) _____
 Other (specify) _____

<u>Part #</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
6808455	Fuel Cell		1	
MPCAL 71			1	
		Repair Time	.5	(Time \$)

TOTAL REPAIR \$ _____

REPAIR NOTES:
 Replaced fuel cell. Upgraded U.P.I.T TO NS-3.10.
 full QC & OPS CHK WITH CALIBRATION.

DATE 1905 04 Technician TC
 Signature _____