

RETURN AND REPAIR FORM



Customer Information

B: _____

S: _____

Company Name: N.J. State Police, West Trenton, NJ

Date Received: 01/29/14

Date given to service: 01/29/14

Carrier: FedEx UPS USPS Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR NL-0046

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: June 2015

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information

Test#: _____

Part Number	Description	Qty	Total Cost
MP Cal 71	Calibration	1	NC/W
MP Labor	Labor	10	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W
8312928	Display	1	NC/W
12029	Power Supply	1	NC/W

Repair Notes: Replaced fuel cell, spacer plates, display, and power supply.

Display had some faded segments in it and power supply was causing the unit to not power on properly.

Tested the unit over the course of a few days to make sure the issue the customer reported with display blinking was resolved.

Fuel Cell = AREL-1091

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 02/17/14