

Return & Repair Form:

Customer Information:

Company Name: NOSP/000 B: _____ S: _____
 Date Received: 11-19-08 Date Given to Service: _____
 Carrier: FedEx - UPS - DHL - USPS Method: GRD - NDA - 2DY - 3DY - Other

Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR NL - 0047
 9510- Drug Tester 5000
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -
 Whole Instrument Top 1/2 Sim Ser #: _____
 Other _____ Probe Ser # DD _____ P -
ACCESSORIES:
 110 V A/C Adapter Regulator Mag Card Rdr# _____
 Printer Paper Printer Ribbon Casio # _____
 Mouthpieces Carrying Case Dry Gas
 Other (Please Specify) _____ Warranty Exp. Date _____

Repair Information:

Test #

Reason for Return: FC

| Part Number | Description | Qty | Total Cost |
|------------------|------------------------|-----------|-------------------|
| <u>6808455</u> | <u>FC</u> | <u>1</u> | <u>N/C w/warr</u> |
| <u>6808484</u> | <u>Plate</u> | <u>1</u> | <u>N/C w/warr</u> |
| <u>MP Cal 71</u> | <u>Cal</u> | <u>1</u> | <u>N/C w/warr</u> |
| <u>MISC Part</u> | <u>Pressure Sensor</u> | <u>1</u> | <u>N/C w/warr</u> |
| <u>MP Labor</u> | <u>Labor</u> | <u>.5</u> | <u>N/C w/warr</u> |

Repair Notes:

FC ARAE-0801 replaced with FC ARAJ-0027 Date AUG, 2009
Cal w/ QC & Ops Check
Re-claimed Part Pressure Sensor

Service Technician: JS

Date: 12/01/09