

RETURN & REPAIR FORM



CUSTOMER NJ State Police Cust. # _____
res/pt 612

Date ret. 5/25 Carrier UPS Method 611 RMA# _____

7110 Serial# ARNL-0048 Warranty exp.: _____

Reported Problem: _____

- Accessories (check all that apply): Regulator P.Paper Scotty V Printer Rib. Pouch
 Sim. Temp. Probe Ser.# _____ Keyboard Organizer Key
 Mouthpieces _____ Sim. To cuvette hose Pump to Sim. hose Power cord
 Other (specify) _____
 Other (specify) _____

<u>Part#</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
<u>6908455</u>	<u>Fuel Cell</u>		<u>1</u>	
<u>MPCAL71</u>			<u>1</u>	
		Repair Time	<u>.5</u>	Time \$

TOTAL REPAIR \$ _____

REPAIR NOTES:

Upgraded Unit to NS-3.10. Calibrated Unit
full per OPS CHC. Replaced Fuel Cell.

18 OCT 04 Edmann

Company name and address (Manufacturer) _____