


RETURN & REPAIR FORM



CUSTOMER NJ State Police Cust. # 703-226 612

Date ret. 5/25 Carrier UPS Method G1J RMA#

7110 Serial# ARAL-0049 Warranty exp.:

Reported Problem:

- Accessories (check all that apply): ☐ Regulator ☐ P.Paper ☐ Scotty V ☐ Printer Rib. ☐ Pouch
 Sim. Temp. Probe Ser.# ☐ Keyboard ☐ Organizer ☐ Key
☐ Mouthpieces ☐ Sim. To cuvette hose ☐ Pump to Sim. hose ☐ Power cord
☐ Other (specify)
☐ Other (specify)

<u>Part #</u>	<u>Description</u>	<u>\$</u>	<u>Qty</u>	<u>Total</u>
6408455	Fuel Cell		1	
MPCAL 71			1	
		Repair Time	.5	Time \$

TOTAL REPAIR \$

REPAIR NOTES:

Upgraded Unit to NS-3.10 Replaced Fuel Cell
 full @ 4 OPS CKE WITH Calibration

Date 10/18/04 Technician EW

Customer's receipt is included ☐ Billing