

# Return & Repair Form:

## Customer Information:

Company Name: NOSP/DOO B: \_\_\_\_\_ S: \_\_\_\_\_  
 Date Received: 11-19-08 Date Given to Service: \_\_\_\_\_  
 Carrier: FedEx - UPS - DHL - USPS Method: GBD - NDA - 2DY - 3DY - Other

## Product Information:

Product: 7410-7110-6510-6810-7510 Serial #: AR NL - 0049  
 9510- Drug Tester 5000  
 Description: A - B - Plus - Screener - Demo Printer Ser #: AR -  
 Whole Instrument Top 1/2 Sim Ser #: \_\_\_\_\_  
 Other \_\_\_\_\_ Probe Ser # DD P -  
**ACCESSORIES:**  
 110 V A/C Adapter  Regulator  Mag Card Rdr# \_\_\_\_\_  
 Printer Paper  Printer Ribbon  Casio # \_\_\_\_\_  
 Mouthpieces  Carrying Case  Dry Gas  
 Other (Please Specify) \_\_\_\_\_ Warranty Exp. Date \_\_\_\_\_

## Repair Information:

FL

Test #

Reason for Return: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>6809455</u>	<u>FL</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Cal 71</u>	<u>Cal</u>	<u>1</u>	<u>N/C WARR</u>
<u>MP Labor</u>	<u>Labor</u>	<u>.5</u>	<u>N/C WARR</u>

### Repair Notes:

NEW FL ARAJ-0233  
Cal w/DC & Ops Check

Service Technician: JS

Date: 11/25/09