

RETURN & REPAIR FORM



CUSTOMER NJ State Police Cust. # _____
 Date ret. 5/25 Carrier YPS Method G-1 RMA# _____
 7110 Serial# ARNL-0050 Warranty exp.: _____

Reported Problem: _____

- Accessories (check all that apply): Regulator P.Paper Scotty V Printer Rib. Pouch
 Sim. Temp. Probe Ser.# _____ Keyboard Organizer Key
 Mouthpieces _____ Sim. To cuvette hose Pump to Sim. hose Power cord
 Other (specify) _____
 Other (specify) _____

<u>Part#</u>	<u>Description</u>	<u>S</u>	<u>Qty</u>	<u>Total</u>
6908455	Fuel Cell		1	
MPCAL 41			1	
		Repair Time	.5	Time \$

TOTAL REPAIR \$ _____

REPAIR NOTES:
 Replaced Fuel Cell. Calibrated Unit. Full O2 &
 O2S CHL. WITH firmware Upgrade to NS-3.10.

Date 19 Oct 04 Technician [Signature]
 I certify that the item is ready for return Shipping